



Monobind Inc.

The World Resource for Diagnostic Products

VALUE ASSIGNMENT FOR BIO-RAD  
MASTER LOT 6493(L-3) EXP:2023/9/30

AccuLite® CLIA	6493L		64931		64932		64933	
	Mean	Range	Mean	Range	Mean	Range	Mean	Range
Liquicheck Specialty IA								
anti-TG in IU/ml	52.55	35.21-69.89	112.20	75.17-149.23	196.21	131.46-260.96	356.37	238.77-473.97
anti-TPO in IU/ml	9.93	6.65-13.20	20.20	13.53-26.87	41.09	27.53-54.64	113.97	76.36-151.58
C-peptide in ng/ml	0.69	0.46-0.91	0.73	0.49-0.97	2.88	1.93-3.83	8.22	5.51-10.93

DOCUMENT HISTORY

PREPARED BY: [Signature] DEPT: QC  
 APPROVED BY: [Signature] DEPT: Administration  
 REVISION: 0

VERIFIED BY: [Signature] DEPT: QA  
 EFFECTIVE DATE: 2020-10-14  
 DCO: N/A