

**VALUE ASSIGNMENT FOR BIO-RAD  
MASTER LOT 4229(1-3) EXP:2024/04/30**

AccuBind® ELISA	42291		42292		42293	
	Mean	Range	Mean	Range	Mean	Range
<b>Fertility</b>						
Ferritin in ng/ml	17.66	11.83-23.48	52.28	35.02-69.53	103.01	69.02-137.00
AFP in ng/ml	23.59	15.81-31.37	88.85	59.53-118.16	362.21	242.68-481.74
LH in mIU/ml	154.66	103.62-205.70	57.13	38.28-75.99	38.94	26.09-51.79
hCG in mIU/ml	<2		43.2	28.95-57.46	23928.66	16032.20-31825.12
Testosterone in ng/ml	0.24	0.16-0.32	3.31	2.22-4.41	7.62	5.10-10.13
Progesterone in ng/ml	23.26	15.59-30.94	43.01	28.82-57.20	>60	
E2 in pg/ml	184.93	123.91-245.96	200.48	134.32-266.64	1039.08	696.18-1381.97

DOCUMENT HISTORY

PREPARED BY: MSW DEPT: QC  
 APPROVED BY: Fred Lopez DEPT: Administration  
 REVISION: 0

VERIFIED BY: Ashatola DEPT: QA  
 EFFECTIVE DATE: 2024-04-15  
 DCO: N/A