

VALUE ASSIGNMENT FOR BIO-RAD
MASTER LOT 6766(1-3)
Exp: 2024/03/31

AccuLite® CLIA	67661		67662		67663	
	Mean	Range	Mean	Range	Mean	Range
Cardiac Markers						
CKMB in ng/ml	3.96	2.65-5.27	15.19	10.18-20.20	70.65	47.34-93.96
hsCRP in µg/ml	1.23	0.82-1.63	2.53	1.69-3.36	7.39	4.95-9.83
Myoglobin in ng/ml	54.36	36.42-72.30	104.36	69.92-138.80	214.74	143.88-285.61
Troponin I in ng/ml	<1		1.34	0.90-1.78	4.21	2.82-5.60

Note: Monobind evaluates levels 1, 2 and 3 in the Bio-Rad control.

It does not evaluate levels 1A, 1B and 1C due to minimal difference in value recognition.

DOCUMENT HISTORY

 PREPARED BY: MSW DEPT: QC

 VERIFIED BY: Ashatola DEPT: QA

 APPROVED BY: [Signature] DEPT: Administration

EFFECTIVE DATE: 2021-05-17

REVISION: 0

DCO: N/A