

VALUE ASSIGNMENT FOR BIO-RAD
MASTER LOT 7460(1-3) Exp:2024/05/28

AccuLite® CLIA	74601		74602		74603	
	Mean	Range	Mean	Range	Mean	Range
Liquicheck Tumor Markers						
Ferritin in ng/ml	58.32	39.08-77.57	283.5	189.94-377.05	594.99	398.64-791.33
CEA in ng/ml	3.25	2.17-4.32	24.77	16.60-32.95	99.07	66.37-131.76
AFP in ng/ml	8.78	5.88-11.67	86	57.62-114.38	261.13	174.95-347.30
PSA in ng/ml	<0.5		6.47	4.33-8.60	32.56	21.81-43.30
fPSA in ng/ml	<0.3		2.89	1.93-3.84	>10	
CA-125 in U/ml	16.49	11.05-21.93	54.33	36.40-72.25	180.64	121.03-240.25
CA-19-9 in U/ml	14.28	9.57-18.99	26.74	17.91-35.56	87.7	58.76-116.64
CEA Next Generation in ng/ml	<5		21.87	14.65-29.08	66.48	44.54-88.41
PRL in ng/ml	14.71	9.85-19.56	46.42	31.10-61.74	>100	
HCG in mIU/ml	4.84	3.24-6.43	60.88	40.79-80.97	>200	
TG in ng/ml	6.62	4.23-9.0	40.98	27.46-54.50	90.28	60.49-120.07
CA 15-3 in U/ml	40.01	26.81-53.21	80.9	54.20-107.60	176.34	118.15-234.53
HCG-XR in mIU/ml	<5		67.35	45.12-89.58	294.71	197.46-391.97
Calcitonin pg/ml	13.87	9.30-18.45	101.52	68.02-135.03	479.43	321.22-637.64
Aldosterone pg/ml	125.43	84.04-166.82	384.06	257.32-510.80	847.64	567.92-1127.36

DOCUMENT HISTORY

 PREPARED BY: RSW DEPT: QC

 VERIFIED BY: Ashatola DEPT: QA

 APPROVED BY: Fred Lopez DEPT: Administration

EFFECTIVE DATE: 2021-06-03

REVISION: 0

DCO: N/A