



VALUE ASSIGNMENT FOR BIO-RAD
MASTER LOT 8783(1-3) Exp:03/31/2024

AccuLite® CLIA	87831		87832		87833	
	Mean	Range	Mean	Range	Mean	Range
Cardiac Markers						
CKMB in ng/ml	2.09	1.25-2.94	12.68	8.49-16.86	65.24	43.71-86.77
hsCRP in µg/ml	0.62	0.35-0.89	2.36	1.58-3.13	6.02	3.32-8.72
Troponin in ng/ml	<0.5		2.02	1.35-2.68	5.26	3.53-7.00
Myoglobin in ng/ml	36.12	24.20-48.04	125.64	84.18-167.11	220.56	147.78-293.35

DOCUMENT HISTORY

PREPARED BY: NSW DEPT: QC VERIFIED BY: Ashatola DEPT: QA
APPROVED BY: Fred Lopez DEPT: Administration EFFECTIVE DATE: 2021-JUN-18
REVISION: 0 DCO: N/A