

**VALUE ASSIGNMENT FOR BIO-RAD
MASTER LOT 6494(L-3) EXP:2024/4/30**

AccuLite® CLIA	6494L		64941		64942		64943	
	Mean	Range	Mean	Range	Mean	Range	Mean	Range
Liquicheck Specialty IA								
anti-TG in IU/ml	47.03	31.51-62.55	98.89	66.26-131.53	160.77	107.71-213.82	348.08	233.22-462.95
anti-TPO in IU/ml	8.98	6.01-11.94	20.53	13.76-27.31	48.95	32.80-65.10	119.44	80.02-158.85
C-peptide in ng/ml	0.75	0.50-1.0	0.74	0.35-1.14	2.68	1.80-3.56	7.61	5.10-10.12

DOCUMENT HISTORY

PREPARED BY: *MSW* DEPT: QC
 APPROVED BY: *Falga* DEPT: Administration
 REVISION: 0

VERIFIED BY: *Ashata* DEPT: QA
 EFFECTIVE DATE: 2021-07-01
 DCO: N/A