

VALUE ASSIGNMENT FOR BIO-RAD						
MASTER LOT 7461(1-3)			Exp:2024/09/30			
AccuLite® CLIA	74611		74612		74613	
	Mean	Range	Mean	Range	Mean	Range
Lyphocheck Tumor Markers						
Ferritin in ng/ml	48.61	32.57-64.65	303.1	203.07-403.12	>700	
CEA in ng/ml	2.78	1.86-3.69	23.06	15.45-30.67	82	54.94-109.06
AFP in ng/ml	8.12	5.44-10.80	84.12	56.36-111.88	263.09	176.27-349.91
PSA in ng/ml	0.18	0.12-0.24	6.14	4.11-8.17	27.99	15.59-40.38
fPSA in ng/ml	<0.3		2.88	1.93-3.83	>10	
CA-125 in U/ml	19.31	12.94-25.69	63.99	42.87-85.10	210.81	141.24-280.38
CA-19-9 in U/ml	14.76	9.89-19.63	27.98	17.34-38.61	86.17	57.73-114.61
CEA Next Generation in ng/ml	<5		25.09	16.81-33.37	85.13	57.04-113.22
PRL in ng/ml	14.8	9.92-19.68	53.13	35.59-70.66	>100	
HCG in mIU/ml	4.94	3.31-6.58	74.26	40.06-108.45	>200	
TG in ng/ml	5.28	3.54-7.03	27.95	17.45-38.46	84.06	56.32-111.81
CA 15-3 in U/ml	37.24	24.95-49.52	81.9	54.87-108.92	170.23	114.06-226.41
HCG-XR in mIU/ml	4.38	2.94-5.83	61.73	41.36-82.10	313.71	210.19-417.24
Calcitonin pg/ml	9.21	5.72-12.71	83.62	56.02-111.21	378.07	253.31-502.83
Aldosterone pg/ml	130.18	87.22-173.14	311.54	208.73-414.35	1062.86	712.11-1413.60

DOCUMENT HISTORY			
PREPARED BY: <u>NSW</u>	DEPT: QC	VERIFIED BY: <u>AShatola</u>	DEPT: QA
APPROVED BY: <u>Fred Lopez</u>	DEPT: Administration	EFFECTIVE DATE: 2021-10-19	
REVISION: 0		DCO: N/A	