

**VALUE ASSIGNMENT FOR BIO-RAD**
**MASTER LOT 6767(1-3)**
**Exp: 2024/08/31**

AccuLite® CLIA	67671		67672		67673	
	Mean	Range	Mean	Range	Mean	Range
Liquicheck Cardiac Markers LT Plus						
CKMB in ng/ml	4.48	3-5.96	20.6	13.80-27.40	101.12	67.75-134.49
hsCRP in µg/ml	1.17	0.78-1.55	2.52	1.69-3.36	6.76	3.80-9.71
Myoglobin in ng/ml	73.87	49.50-98.25	96.91	64.93-128.88	203.6	136.41-270.79
Troponin I in ng/ml	<0.5		1.04	0.62-1.46	4.18	2.80-5.56

**Note: Monobind evaluates levels 1, 2 and 3 in the Bio-Rad control.**

**It does not evaluate levels 1A, 1B and 1C due to minimal difference in value recognition.**

## DOCUMENT HISTORY

 PREPARED BY: *MSW* DEPT: QC

 VERIFIED BY: *Ashatola* DEPT: QA

 APPROVED BY: *Fred Lopez* DEPT: Administration

EFFECTIVE DATE: 2021-11-02

REVISION: 0

DCO: N/A