

**VALUE ASSIGNMENT FOR BIO-RAD
MASTER LOT 6497(L,1,2,3) EXP:2025-06-30**

AccuLite® CLIA	6497L		64971		64972		64973	
	Mean	Range	Mean	Range	Mean	Range	Mean	Range
Liquicheck Specialty IA								
anti-TG in IU/ml	42.93	28.76-57.09	94.10	63.05-125.15	168.28	112.75-223.81	342.03	229.16-454.89
anti-TPO in IU/ml	10.54	7.06-14.02	22.58	15.13-30.03	61.26	41.04-81.48	138.70	92.93-184.47
C-peptide in ng/ml	0.89	0.59-1.18	0.85	0.57-1.13	3.44	2.31-4.58	>9	
Vit-D Direct in ng/ml	22.12	14.82-29.41	24.16	16.19-32.14	45.68	30.60-60.75	88.33	59.18-117.48

DOCUMENT HISTORY

PREPARED BY: *NSW* DEPT: QC
 APPROVED BY: *Fred Lopez* DEPT: Administration
 REVISION: 0

VERIFIED BY: *Ashatola* DEPT: QA
 EFFECTIVE DATE: 2022-08-08
 DCO: N/A

