

VALUE ASSIGNMENT FOR BIO-RAD						
MASTER LOT4859(1-3)			Exp:2024/09/30			
AccuLite® CLIA	48591		48592		48593	
	Mean	Range	Mean	Range	Mean	Range
Liquicheck AMH Control						
AMH in ng/ml	1.11	0.75-1.48	8.24	4.45-12.03	>15	

DOCUMENT HISTORY			
PREPARED BY: <u>NSW</u>	DEPT: QC	VERIFIED BY: <u>A Shatola</u>	DEPT: QA
APPROVED BY: <u>Fred Lopez</u>	DEPT: Administration	EFFECTIVE DATE: 2022-09-18	
REVISION: 0		DCO: N/A	