

VALUE ASSIGNMENT FOR BIO-RAD						
MASTER LOT4859(1-3)			Exp:2024/09/30			
	48591		48592		48593	
AccuBind® ELISA	Mean	Range	Mean	Range	Mean	Range
Liquicheck AMH Control						
AMH in ng/ml	1.21	0.81-1.61	8.47	5.68-11.27	>15	

DOCUMENT HISTORY

PREPARED BY: \_\_\_\_\_\_\_DEPT: QC

APPROVED BY: \_\_\_\_\_\_\_DEPT: Administration AShatola DEPT: QA VERIFIED BY:\_

REVISION: 0

DCO: N/A

EFFECTIVE DATE:

2022-09-18