

VALUE ASSIGNMENT FOR BIO-RAD						
MASTER LOT 9494(1-3)			Exp:2024/10/31			
AccuBind® ELISA	94941		94942		94943	
	Mean	Range	Mean	Range	Mean	Range
Liquicheck Tumor Markers						
Ferritin in ng/ml	29.97	20.08-39.86	310.42	207.98-412.86	643.26	430.99->800
CEA in ng/ml	3.6	2.41-4.79	30.01	20.11-39.92	106.46	71.33-141.60
AFP in ng/ml	6.74	4.41-9.06	85.71	57.43-114.0	263.96	176.35-351.06
PSA in ng/ml	<0.5		10.5	7.04-13.97	46.96	31.46-62.46
fPSA in ng/ml	<0.5		4.09	2.74-5.44	>10	
CA-125 in U/ml	20.22	13.55-26.90	60.52	40.55-80.49	193.34	129.54-257.14
CA-19-9 in U/ml	5.49	3.68-7.30	23.36	15.65-31.07	69.36	46.47-92.25
CEA Next Generation in ng/ml	<5		26.14	17.51-34.76	94.46	63.29-125.63
PRL in ng/ml	12.28	8.23-16.34	38.88	26.05-51.71	>100	
PRL in ng/ml-human	10.46	7.01-13.91	31.25	20.93-41.56	>100	
HCG in mIU/ml	5.5	3.68-7.31	56.44	37.82-75.07	>200	
TG in ng/ml	4.42	2.96-5.88	47.92	32.10-63.73	117.42	78.67-156.17
CA 15-3 in U/ml	19.47	13.04-25.89	43.27	28.99-57.55	85.85	57.52-114.17
HCG-XR in mIU/ml	3.86	2.58-5.13	52.88	35.43-70.33	210.58	141.09-280.07

DOCUMENT HISTORY			
PREPARED BY: <u>MSW</u>	DEPT: QC	VERIFIED BY: <u>Ashatola</u>	DEPT: QA
APPROVED BY: <u>Fred Lopez</u>	DEPT: Administration	EFFECTIVE DATE:	2021-10-26
REVISION: 0		DCO: N/A	