

VALUE ASSIGNMENT FOR BIO-RAD						
MASTER LOT 9494(1-3)			Exp:2024/10/31			
AccuLite® CLIA	94941		94942		94943	
	Mean	Range	Mean	Range	Mean	Range
Liquicheck Tumor Markers						
Ferritin in ng/ml	30.48	20.42-40.53	325.6	188.60-462.59	696.01	332.30->800
CEA in ng/ml	3.3	2.21-4.39	28.4	19.03-37.77	92.6	62.04-123.16
AFP in ng/ml	7	4.69-9.31	95.1	63.72-126.48	259.7	174.0-345.40
PSA in ng/ml	<5		10.63	7.12-14.14	49.59	33.23-65.95
fPSA in ng/ml	<5		4	2.68-5.32	>10	
CA-125 in U/ml	18.27	12.24-24.29	56.72	38.0-75.43	175.01	117.26-232.76
CA-19-9 in U/ml	4.15	2.78-5.52	21.13	14.15-28.10	66.23	44.37-88.09
CEA Next Generation in ng/ml	<5		26.33	17.64-35.02	86.78	58.14-115.41
PRL in ng/ml	11.25	7.54-14.96	37.6	25.19-50.01	>100	
HCG in mIU/ml	5.58	3.74-7.41	55.9	37.45-74.35	237.9	159.39->250
TG in ng/ml	5	3.35-6.65	44.83	30.03-59.63	112.48	75.38-149.59
CA 15-3 in U/ml	18.15	12.16-24.13	41.18	27.59-54.77	78.86	52.83-104.88
HCG-XR in mIU/ml	3.9	2.61-5.19	52.17	34.95-69.38	211.53	141.73-281.34

DOCUMENT HISTORY			
PREPARED BY: <u>NSW</u>	DEPT: QC	VERIFIED BY: <u>Ashatoka</u>	DEPT: QA
APPROVED BY: <u>Fuelgum</u>	DEPT: Administration	EFFECTIVE DATE:	2021-10-26
REVISION: 0		DCO: N/A	