

VALUE ASSIGNMENT FOR BIO-RAD						
MASTER LOT 9495(1-3)			Exp:2025/02/28			
AccuLite® CLIA	94951		94952		94953	
	Mean	Range	Mean	Range	Mean	Range
Liquicheck Tumor Markers						
Ferritin in ng/ml	28.81	19.30-38.32	320.47	214.71-426.22	634.5	425.11->800
CEA in ng/ml	3.1	2.08-4.12	29.9	20.03-39.77	106.2	71.15-141.25
AFP in ng/ml	5.83	3.90-7.75	93.65	62.75-124.55	286.9	192.22-381.58
PSA in ng/ml	<2.0		8.43	5.64-11.21	39.15	26.23-52.07
fPSA in ng/ml	None reported					
CA-125 in U/ml	20.84	13.97-27.72	58.99	39.52-78.46	186.73	125.11-248.35
CA-19-9 in U/ml	6.27	4.20-8.33	27.32	18.30-36.34	84.21	56.42-111.99
CEA Next Generation in ng/ml	<5		25.98	17.40-34.55	85.72	57.43-114.01
PRL in ng/ml	9.1	6.10-12.10	33.18	22.23-44.12	>100	
HCG in mIU/ml	5.33	3.57-7.08	57.05	38.22-75.88	>200	
TG in ng/ml	4.5	3.02-5.99	37.6	25.19-50.01	124.3	83.28-165.32
CA 15-3 in U/ml	18.34	12.28-24.39	49.72	33.31-66.13	96.94	64.95-128.92
HCG-XR in mIU/ml	2.18	1.46-2.89	54.07	36.22-71.91	244.07	163.52-324.61

DOCUMENT HISTORY			
PREPARED BY: <u>NSW</u>	DEPT: QC	VERIFIED BY: <u>Ashatoka</u>	DEPT: QA
APPROVED BY: <u>Fred Lopez</u>	DEPT: Administration	EFFECTIVE DATE:	2022/12/01
REVISION: 0		DCO: N/A	