

VALUE ASSIGNMENT FOR BIO-RAD						
MASTER LOT 7462(1-3)			Exp:2025/05/31			
AccuLite® CLIA	74621		74622		74623	
	Mean	Range	Mean	Range	Mean	Range
Lyphocheck Tumor Markers						
Ferritin in ng/ml	52.74	35.34-70.15	308.58	206.75-410.41	>700	
CEA in ng/ml	3.08	2.06-4.09	27.68	18.54-36.81	97.63	65.41-129.85
AFP in ng/ml	6.63	4.44-8.82	79.52	53.28-105.76	234.65	157.22-312.08
PSA in ng/ml	0.70	0.47-0.93	9.6	6.43-12.77	42.58	28.53-56.62
fPSA in ng/ml	<0.3		5.05	3.39-6.72	>10	
CA-125 in U/ml	18.7	12.53-24.87	60.12	40.28-79.95	212.38	142.29-282.46
CA-19-9 in U/ml	20.52	13.75-27.29	37.4	25.06-49.74	101.57	61.25-141.89
CEA Next Generation in ng/ml	<5		22.15	14.84-29.46	87.54	58.65-116.43
PRL in ng/ml	12.74	8.53-16.94	41.63	27.89-55.36	>100	
HCG in mIU/ml	7.25	4.86-9.64	89.25	59.80-118.70	>200	
TG in ng/ml	6.91	4.63-9.19	36	24.12-47.88	79.25	53.10-105.40
CA 15-3 in U/ml	34.51	23.12-45.90	74.16	49.69-98.63	156.74	105.02-208.47
HCG-XR in mIU/ml	5.2	3.48-6.92	73.43	49.20-97.67	332.07	222.48-441.65
Calcitonin pg/ml	8.12	5.44-10.79	68.89	46.16-91.62	325.82	218.30-433.34
Aldosterone pg/ml	151.87	101.75-201.98	419.8	281.27-558.33	1045.43	700.44-1390.43
ACTH pg/ml	59.7	40.0-79.4	236.19	158.25-314.13	597.04	400.02-794.06

DOCUMENT HISTORY			
PREPARED BY: <u>NSW</u>	DEPT: QC	VERIFIED BY: <u>Ashatola</u>	DEPT: QA
APPROVED BY: <u>Fred Lopez</u>	DEPT: Administration	EFFECTIVE DATE: 2022-06-16	
REVISION: 0		DCO: N/A	