

VALUE ASSIGNMENT FOR BIO-RAD						
MASTER LOT 6770(1-3)			Exp: 2026/02/28			
AccuLite® CLIA	67701		67701		67701	
	Mean	Range	Mean	Range	Mean	Range
Liquicheck Cardiac Markers LT Plus						
CKMB in ng/ml	3.39	2.27-4.51	15.45	10.35-20.55	85.82	57.50-114.14
hsCRP in µg/ml	0.8	0.53-1.06	2.08	1.40-2.77	6.41	4.29-8.52
Myoglobin in ng/ml	66.85	44.79-88.91	131.55	88.14-174.96	258.7	173.33-344.07
Troponin I in ng/ml	<0.5		1.92	1.29-2.56	6.33	4.24-8.41

Note: Monobind evaluates levels 1, 2 and 3 in the Bio-Rad control.

It does not evaluate levels 1A, 1B and 1C due to minimal difference in value recognition.

DOCUMENT HISTORY

PREPARED BY: *MSW* DEPT: QC VERIFIED BY: *Ashatola* DEPT: QA
 APPROVED BY: *Fred Lopez* DEPT: Administration EFFECTIVE DATE: 2023-01-21
 REVISION: 0 DCO: N/A