



VALUE ASSIGNMENT FOR BIO-RAD

MASTER LOT 7443(1.2.3.L)

Exp:8/31/2026

AccuLite® CLIA	74431		74432		74433		7443L	
	Mean	Range	Mean	Range	Mean	Range	Mean	Range
Liquicheck D-Dimer								
D-Dimer in ng/ml	369.19	247.36-491.02	1458.94	977.49-1940.38	3050.24	2043.66-4056.82	250.98	168.16-333.81

DOCUMENT HISTORY

PREPARED BY: MSW DEPT: QC

VERIFIED BY: AShatola DEPT: QA

APPROVED BY: Fred Lopez DEPT: Administration

EFFECTIVE DATE:2023/05/24

REVISION: 0

DCO: N/A