

VALUE ASSIGNMENT FOR BIO-RAD								
MASTER LOT 7443(1.2.3.L) Exp:8/31/2026								
AccuBind® ELISA	74431		74432		74433		7443L	
	Mean	Range	Mean	Range	Mean	Range	Mean	Range
Liquicheck D-Dimer								
D-Dimer in ng/ml	393.46	263.62-523.31	1441.92	966.09-1917.75	3159.57	2116.91-4202.22	237.56	159.17-315.96

DOCUMENT HISTORY

PREPARED BY: *MSW* DEPT:QC
 APPROVED BY: *Fred Lopez* DEPT: Administration
 REVISION: 0

VERIFIED BY: *A Shatola* DEPT: QA
 EFFECTIVE DATE:2023/05/24
 DCO: N/A