

VALUE ASSIGNMENT FOR BIO-RAD						
	MASTER	LOT 6771(1	-3) Exp: 2026/06/30			
	67711		67711		67711	
AccuLite® CLIA	Mean	Range	Mean	Range	Mean	Range
Liquicheck Cardiac Markers LT Plus						
CKMB in ng/ml	4.03	2.70-5.36	14.1	9.45-18.75	74.53	49.93-99.12
hsCRP in µg/ml	0.8	0.53-1.06	2.25	1.50-2.99	8.13	5.45-10.81
Myoglobin in ng/ml	74.28	49.76-98.79	126.65	84.19-167.11	260	174.20-345.8
Troponin I in ng/ml	<0.5		1.42	0.95-1.88	5.44	3.64-7.24

Note: Monobind evaluates levels 1, 2 and 3 in the Bio-Rad control. It does not evaluate levels 1A, 1B and 1C due to minimal difference in value recognition.