

VALUE ASSIGNMENT FOR BIO-RAD
MASTER LOT 7464(1-3) Exp:2026/08/31

AccuLite® CLIA	74641		74642		74643	
	Mean	Range	Mean	Range	Mean	Range
Lyphocheck Tumor Markers						
Ferritin in ng/ml	47.41	31.77-63.06	251.79	168.70-334.88	548.31	367.37-729.25
CEA in ng/ml	2.62	1.62-3.61	23.24	15.57-30.91	77.38	51.85-102.92
AFP in ng/ml	5.59	3.74-7.43	65.34	43.78-86.91	203.36	136.25-270.47
PSA in ng/ml	<0.5		8.22	5.51-10.94	37.08	24.85-49.32
fPSA in ng/ml	<0.3		2.56	1.71-3.40	>10	
CA-125 in U/ml	17.39	11.65-23.13	56.06	37.56-74.56	167.73	112.38-223.09
CA-19-9 in U/ml	12.47	8.35-16.58	32.18	21.56-42.79	80.8	54.14-107.46
CEA Next Generation in ng/ml	<5		23.38	15.66-31.10	76.91	51.53-102.29
PRL in ng/ml	11.35	7.60-15.10	40.78	27.32-54.23	>100	
HCG in mIU/ml	6.24	4.18-8.30	74.71	50.06-99.37	>200	
TG in ng/ml	6.63	4.44-8.82	20.73	13.89-27.58	63.53	42.57-84.50
CA 15-3 in U/ml	28.28	18.95-37.61	57.37	38.43-76.30	95.18	63.77-126.59
HCG-XR in mIU/ml	<5		82.42	55.22-109.61	331.16	221.88-440.44
Aldosterone pg/ml	151.6	101.57-201.63	344.43	230.77- 458.10	850.37	569.75-1130.99

DOCUMENT HISTORY

 PREPARED BY: NSW DEPT: QC

 VERIFIED BY: Ashatola DEPT: QA

 APPROVED BY: Fred Lopez DEPT: Administration

EFFECTIVE DATE: 2023-10-16

REVISION: 0

DCO: N/A