

VALUE ASSIGNMENT FOR BIO-RAD						
MASTER LOT 7465(1-3)			Exp:2026/10/31			
AccuLite® CLIA	74651		74652		74653	
	Mean	Range	Mean	Range	Mean	Range
Lyphocheck Tumor Markers						
Ferritin in ng/ml	51.97	34.82-69.12	356.2	238.65-473.75	855.73	573.34-1138.13
CEA in ng/ml	3.26	2.18-4.34	23.2	15.54-30.86	78.92	52.88-104.96
AFP in ng/ml	6.57	4.40-8.73	73.23	49.06-97.39	209.48	140.35-278.60
PSA in ng/ml	<0.5		7.12	4.77-9.47	32.87	22.02-43.71
fPSA in ng/ml	<0.3		2.9	1.94-3.86	>10	
CA-125 in U/ml	19.47	13.04-25.89	58.22	39.01-77.43	156.63	104.94-208.32
CA-19-9 in U/ml	10.91	7.31-14.50	26.98	18.08-35.88	62.9	42.14-83.66
CEA Next Generation in ng/ml	<5		21.78	14.59-28.97	73.17	49.02-97.31
PRL in ng/ml	12.36	8.28-16.44	37.86	25.36-50.35	>100	
HCG in mIU/ml	3.7	2.48-4.92	69.04	46.26-91.82	>200	
TG in ng/ml	5.5	3.69-7.32	28.7	19.23-38.17	81.56	54.64-108.47
CA 15-3 in U/ml	24.85	16.65-33.05	54.53	36.53-72.52	111.3	74.57-148.03
HCG-XR in mIU/ml	<5		80.36	53.84-106.87	351.57	235.55-467.59
Calcitonin pg/ml	9.32	6.24-12.39	69.31	46.43-92.18	338.23	226.62-449.85
Aldosterone pg/ml	128.8	86.30-171.30	327.38	219.34-435.41	766.08	513.27-1018.88

DOCUMENT HISTORY			
PREPARED BY: <u>NSW</u>	DEPT: QC	VERIFIED BY: <u>AShatola</u>	DEPT: QA
APPROVED BY: <u>Fred Lopez</u>	DEPT: Administration	EFFECTIVE DATE: 2023-10-16	
REVISION: 0		DCO: N/A	