



Monobind Inc.

The World Resource for Diagnostic Products

VALUE ASSIGNMENT FOR BIO-RAD

MASTER LOT 5412(1-3) Exp date: 2025/12/31

AccuBind® ELISA	54121		54122		54123	
	Mean	Range	Mean	Range	Mean	Range
Liquichek Maternal Serum 1st Trimester						
hCG in mIU/mL	2735.40	1802.74-3668.06	12786.20	8566.75-17005.65	30410.60	20375.10-40446.10
fβ-hCG in ng/ml	11.64	7.80-15.48	56.90	38.12-75.68	126.53	84.78-168.29
hCG-XR in mIU/mL	2531.33	1695.99-3366.67	12857.67	8614.64-17100.70	29069.67	19476.68-38662.66
PAPP-A in µg/ml	6.99	4.69-9.30	21.47	14.39-28.56	>32	

Note 1: The HCG data was obtained by diluting Level I , II and III at 1:1000 dilution while the HCG-XR data was obtained by diluting Level I , II and III at 1:100 dilution respectably . The dilution factor was later applied to generate the correct value for each level.

Note 2: PAPP-A data is being reported in µg/ml. To convert µg/ml into mIU/l multiply the values x 156.25

DOCUMENT HISTORY

PREPARED BY: *NYW*
APPROVED BY: *Falalga*
REVISION: 0

DEPT: QC
DEPT: Administration

VERIFIED BY: *Ashatola* DEPT: QA
EFFECTIVE DATE: 2024-01-03
DCO: N/A