

VALUE ASSIGNMENT FOR BIO-RAD
MASTER LOT 8788(1-3) Exp:2026/12/31

AccuLite® CLIA	87881		87882		87883	
	Mean	Range	Mean	Range	Mean	Range
Liquicheck Cardiac Markers LT Plus						
CKMB in ng/ml	2.80	1.88-3.72	15.33	10.14-20.51	71.90	48.17-95.63
hsCRP in µg/ml	0.61	0.41-0.81	2.16	1.45-2.87	7.09	4.75-9.42
Myoglobin in ng/ml	45.40	30.42-60.38	148.08	99.21-196.95	335.68	224.91-446.45
Troponin in ng/ml	0.33	0.22-0.44	2.60	1.68-3.52	5.89	3.95-7.84

DOCUMENT HISTORY

PREPARED BY: *MSW* DEPT: QC
 APPROVED BY: *Fred Lopez* DEPT: Administration
 REVISION: 0

VERIFIED BY: *A Shatola* DEPT: QA
 EFFECTIVE DATE: 2024/03/01
 DCO: N/A