

VALUE ASSIGNMENT FOR BIO-RAD						
MASTER LOT 9498(1-3)			Exp:2026/04/30			
AccuBind® ELISA	94981		94982		94983	
	Mean	Range	Mean	Range	Mean	Range
Liquicheck Tumor Markers						
PRL in ng/ml	10.28	6.89-13.67	33.6	22.51-44.69	>100	
PRL in ng/ml-human	11.74	7.87-15.62	38.78	25.98-51.58	>100	
HCG in mIU/ml	3.79	2.54-5.04	76.48	51.24-101.72	>200	
CEA in ng/ml	3.62	2.43-4.82	29.41	19.70-39.11	110.75	74.20-147.30
AFP in ng/ml	7.83	5.25-10.42	81.06	54.31-107.81	228.06	152.80-303.32
PSA in ng/ml	<2.0		9.87	6.61-13.12	41.54	27.83-55.25
TG in ng/ml	4.18	2.80-5.56	42.47	28.45-56.48	130.83	87.65-174.00
fPSA in ng/ml	<0.5		4.10	2.75-5.45	>10	
Ferritin in ng/ml	29.57	19.81-39.33	335.01	224.46-445.56	663.07	444.25->800
CA-125 in U/ml	22.27	14.92-29.62	59.18	39.65-78.71	163.83	109.76-217.89
CA-19-9 in U/ml	7.8	5.23-10.37	29.77	19.94-39.59	80.24	53.76-106.71
CEA Next Generation in ng/ml	<5		27.35	18.32-36.37	76.48	51.24-101.72
CA 15-3 in U/ml	15.23	10.20-20.26	53.55	35.88-71.21	99.56	66.70-132.41
HCG-XR in mIU/ml	2.86	1.91-3.80	72.91	48.85-96.97	301.68	202.13-401.23

DOCUMENT HISTORY			
PREPARED BY: <u>MSW</u>	DEPT: QC	VERIFIED BY: <u>AShatola</u>	DEPT: QA
APPROVED BY: <u>Fred Lopez</u>	DEPT: Administration	EFFECTIVE DATE: 2024/03/01	
REVISION: 0		DCO: N/A	