

VALUE ASSIGNMENT FOR BIO-RAD						
MASTER LOT 9498(1-3)				Exp:2026/04/30		
AccuLite® CLIA	94981		94982		94983	
	Mean	Range	Mean	Range	Mean	Range
Liquicheck Tumor Markers						
PRL in ng/ml	9.84	6.59-13.09	35.47	23.77-47.18	>100	
HCG in mIU/ml	3.54	2.37-4.71	78.94	52.89-105.00	>200	
CEA in ng/ml	3.3	2.21-4.39	29.54	19.79-39.29	107.14	71.78-142.50
AFP in ng/ml	7.5	5.03-9.98	80.34	53.83-106.85	232.24	155.60-308.88
PSA in ng/ml	<2.0		9.46	6.34-12.58	45.4	30.42-60.38
TG in ng/ml	3.72	2.49-4.94	43	28.81-57.19	132.34	88.67-176.01
fPSA in ng/ml	<0.5		4.06	2.72-5.40	>10	
Ferritin in ng/ml	30.2	20.23-40.17	318.08	213.11-423.05	706.03	473.04->800
CA-125 in U/ml	20.3	13.60-27.0	58.15	38.96-77.34	162.45	108.84-216.06
CA-19-9 in U/ml	6.73	4.51-8.94	30.23	20.25-40.20	82.9	55.54-110.26
CEA Next Generation in ng/ml	<5		24.77	16.59-32.94	76.33	51.14-101.51
CA 15-3 in U/ml	14.6	9.78-19.42	54.52	36.53-72.51	95.01	63.65-126.36
HCG-XR in mIU/ml	3.95	2.65-5.25	66.9	44.82-88.98	318.15	213.16-423.14

DOCUMENT HISTORY			
PREPARED BY: <u>RSW</u>	DEPT: QC	VERIFIED BY: <u>Ashatola</u>	DEPT: QA
APPROVED BY: <u>Fred Lopez</u>	DEPT: Administration	EFFECTIVE DATE: 2024/03/01	
REVISION: 0		DCO: N/A	